

# SAFETY RECALL

## Drive Recalls Bed Assist Handles and Rail

**Hazard:** Users can become entrapped within the bed rail or between the rail and the side of the mattress. This poses a serious entrapment hazard and risk of death by asphyxia.



RTL15063-ADJ (Home Bed Assist Handle)



15064 (Bed Assist Handle)



RTL15062 (Bed Assist Rail w/Folding Board)



RTL15073 (Home Bed Assist Handle)

Store SKU #	Product Description	UPC	Model #	Internet #	Online SKU #
1001412760	Home Bed Assist Handle	822383531588	rtl15073	206196225	1001412760
1000956017	Home Bed Assist Rail and Bed Board Combo	822383126456	15062	205517580	1000956017
1000956018	Home Bed Assist Rail	822383126470	15064	204124187	1000956018
1001477586	Adjustable Height Home Bed Assist Handle	822383246192	RTL15063-ADJ	203472358	1001477586

**Product Description:** The recall involves 3 models of bed assist handles and 1 bed assist rail which were sold between 2007 and December 2021.

**Remedy:** Consumers should immediately stop using the recalled products and contact Drive to receive a refund.

### For More Information

Visit [www.recallrtr.com/bedhandles](http://www.recallrtr.com/bedhandles) or call 1-877-467-3099 from 8 a.m. to 7 p.m. ET, Monday through Friday.